



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005  
ANSI/NCSL Z540-1-1994

ANRITSU COMPANY CALIBRATION SERVICE  
 TX FACILITY  
 1155 East Collins Blvd.  
 Richardson, TX 75081  
 Yeou-Song (Brian) Lee Phone: 408 201 1976

CALIBRATION

Valid To: April 30, 2016

Certificate Number: 2160.02

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1</sup>:

I. Electrical – RF/Microwave

Parameter/Range	Frequency	CMC <sup>2</sup> (±)	Comments
Power Level <sup>3</sup> –  0 dBm	(10 to 50) MHz (50 to 150) MHz (0.15 to 2) GHz (2 to 12) GHz (12 to 18) GHz (18 to 32) GHz (32 to 40) GHz	(0.090 to 0.080) dB + <i>M</i> 0.080 dB + <i>M</i> 0.080 dB + <i>M</i> 0.080 dB + <i>M</i> (0.080 to 0.11) dB + <i>M</i> (0.080 to 0.11) dB + <i>M</i> (0.10 to 0.12) dB + <i>M</i>	Direct power measurement (for Type N and Type K connector), MA 247XA/B with ML 2437/8A and ML 2530A  <i>M</i> = Mismatch
(10 to -60) dBm (except 0 dBm)	(10 to 50) MHz (50 to 150) MHz (0.15 to 2) GHz (2 to 12) GHz (12 to 18) GHz (18 to 32) GHz (32 to 40) GHz	(0.13 to 0.12) dB + <i>M</i> 0.12 dB + <i>M</i> 0.12 dB + <i>M</i> (0.12 to 0.13) dB + <i>M</i> (0.12 to 0.14) dB + <i>M</i> (0.13 to 0.17) dB + <i>M</i> (0.16 to 0.18) dB + <i>M</i>	

Parameter/Range	Frequency	CMC <sup>2</sup> (±)	Comments
Power Level <sup>3</sup> (cont) – (20 to -100) dBm:  (-60 to -85) dBm	(10 to 50) MHz (50 to 150) MHz (0.15 to 2) GHz (2 to 12) GHz (12 to 18) GHz (18 to 32) GHz (32 to 40) GHz	0.13 dB + <i>M</i> 0.13 dB + <i>M</i> 0.13 dB + <i>M</i> 0.13 dB + <i>M</i> (0.13 to 0.14) dB + <i>M</i> (0.14 to 0.17) dB + <i>M</i> (0.17 to 0.18) dB + <i>M</i>	Direct power measurement (for Type N and Type K connector), MA 247XA/B with ML 2437/8A and ML 2530A  <i>M</i> = Mismatch
(-85 to -95) dBm	(10 to 50) MHz (50 to 150) MHz (0.15 to 2) GHz (2 to 12) GHz (12 to 18) GHz (18 to 32) GHz (32 to 40) GHz	0.17 dB + <i>M</i> 0.17 dB + <i>M</i> (0.17 to 0.16) dB + <i>M</i> (0.16 to 0.17) dB + <i>M</i> 0.17 dB + <i>M</i> (0.17 to 0.20) dB + <i>M</i> (0.20 to 0.21) dB + <i>M</i>	
(-95 to -100) dBm	(10 to 50) MHz (50 to 150) MHz (0.15 to 2) GHz (2 to 12) GHz (12 to 18) GHz (18 to 32) GHz (32 to 40) GHz	0.85 dB + <i>M</i> 0.85 dB + <i>M</i> 0.85 dB + <i>M</i> 0.85 dB + <i>M</i> (0.85 to 0.86) dB + <i>M</i> 0.86 dB + <i>M</i> (0.86 to 0.87) dB + <i>M</i>	
Audio Level –  20 Hz to 100 kHz	3 mV to 500 V <sub>p-p</sub>	5.7 % of reading	Peak to peak voltmeter Rohde Schwartz URE3 and Scope Tektronix TDS3052B

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
BER/FER –			
Single Slot BER	(0 to 100) %	0.12 % of reading	Blue Tooth test set MT885X
Single Slot FER	(0 to 100) %	12 % of reading	100 frames/CDMA

Parameter/Equipment	Frequency	CMC <sup>2</sup> (±)	Comments
Error Vector Magnitude (EVM): LTE and WCDMA Format – Measure  Modulation Bandwidth: (0 to 40) MHz	50 Hz to 6 GHz	0.30 % rms of reading	MS2691A

## II. Time & Frequency

Parameter/Equipment	Frequency	CMC <sup>2</sup> (±)	Comments
Frequency – Measuring Equipment	10 MHz	5 parts in 10 <sup>12</sup> Hz/Hz	GPS disciplined oscillator aging rate
Frequency <sup>3</sup> – Measure	10 MHz to 40 GHz	5 × 10 <sup>-12</sup> f	Counter MF241X  f = frequency

<sup>1</sup> This laboratory offers commercial calibration service and field calibration service (where noted).

<sup>2</sup> Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

<sup>3</sup> Field calibration service is available for this calibration and this laboratory meets A2LA R104 – *General Requirements: Accreditation of Field Testing and Field Calibration Laboratories* for these calibrations. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.



American Association for Laboratory Accreditation

# Accredited Laboratory

A2LA has accredited

## ANRITSU COMPANY CALIBRATION SERVICE

*Richardson, TX*

for technical competence in the field of

### Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and any additional program requirements in the field of calibration. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (*refer to joint ISO-ILAC-IAF Communiqué dated 8 January 2009*).

Presented this 20<sup>th</sup> day of May 2014.



  
\_\_\_\_\_  
President & CEO

For the Accreditation Council  
Certificate Number 2160.02  
Valid to April 30, 2016

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*